



**Palacký University in Olomouc**  
**Křížkovského 8, 771 47 OLOMOUC**

No.: 69/PJ/OVZ/2012

Olomouc, February 9, 2012

**Subj.: Sufficient information to Tender documentation No.2**

The tender authority provides sufficient information No. 1 to overlimit tender for supplies called: „**RCPTM – Vacuum steaming equipment**”, published in IS VZ US with the registration number: 7000000003747, according to § 49 of the Act No. 137/2006 of Coll. On Tenders.

**Question No. 1:**

Substrates: max. diameter of substrate is 1000mm, but what about an geometry (mirror, plane plate)?

**Answer:**

The mirror can have various value of the radius of curvature – from infinitum up to 2000 mm.

**Question No. 2**

The substrate heating: until which temperature?

**Answer:**

An infrared heating has to be able to make the substrate heating until the temperature 320°C.

Kind regards

UNIVERZITA PALACKÉHO v OLOMOUCI  
Křížkovského 8, 771 47 Olomouc

Mgr. Petra Jungová  
contact person for tender matters